



# EXPLOSION PROTECTION CERTIFICATE OF CONFORMITY

Cert No. GYJ05241

This is to certify that the product

Compact Electromagnetic Flowmeter

manufactured by Krohne Altometer

(Address: Kerkeplaat 12, 3313LC Dordrecht, the Netherlands)

which model is Optiflux 5300C/...-Ex

Ex marking Ex dme ia(ia) II CT3~T6 / Ex de ia(ia) II CT3~T6

product standard /

drawing number 8.30698.01A、8.30698.02A

has been inspected and certified by NEPSI, and that it conforms

to GB3836.1~4-2000、GB3836.9-1990

This Approval shall remain in force until 2010.08.14

Remarks The notes for safe use see the attachment I.

Director



National Supervision and Inspection Centre for  
Explosion Protection and Safety of Instrumentation

Issued Date 2005.08.15

This Certificate is valid for products compatible with the documents and samples approved by NEPSI.